



## VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the product designated below.

Product	Switching power supply
Model	PMP105x-y (x=Blank,F or SF, y=12,12-1,13,13-1,13-3,14,15,16,17 and 18) ; FSP105-KxAM1 (x=H, G, B, C, A, K, L, E and F)
Brand Name	PROTEK POWER ; FSP
Applicant	<b>FSP Group Inc.</b> 2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C.
Manufacturer	<b>(1) FSP Technology Inc. Kaohsiung Branch. 2-3, East 3rd Street, N.E.P.Z.,</b> P.O. Box 35-25, Kaohsiung Taiwan, R.O.C. <b>(2) Protek Electronic (China) Corp.</b> No. 3, Yong Tai 1st Street , North Area Industry Tian Mei Huang Jiang, Dongguan, Guangdong 523752 China <b>(3) Zhonghan Electronics (Shenzhen) Co., Ltd.</b> Building A2 in Section A, Building 4,6,7 in Section C, JuYuan Industrial Zone, TangWei Village,FuYong Town, BaoAn Dist., ShenZhen City, Guangdong 518103 China
Applicable Standard(s)	<b>EN 55011: 2009 + A1: 2010 (Class A)</b> <b>CISPR 11: 2011</b> EN 61000-3-2: 2014; IEC 61000-3-2: 2009 EN 61000-3-3: 2013; IEC 61000-3-3: 2013 <b>EN 60601-1-2: 2007 + AC: 2010</b> <b>IEC 60601-1-2: 2014 (Edition 4)</b> IEC 61000-4-2: 2008; EN 61000-4-2: 2009 IEC 61000-4-3: 2010; EN 61000-4-3: 2006 + A1: 2008 + A2: 2010 IEC 61000-4-4: 2012; EN 61000-4-4: 2012 IEC 61000-4-5: 2014; EN 61000-4-5: 2014 IEC 61000-4-6: 2013; EN 61000-4-6: 2014 IEC 61000-4-8: 2009; EN 61000-4-8: 2010 IEC 61000-4-11: 2004; EN 61000-4-11: 2004
Reference No.	T151216N09-E
Test Laboratory	<b>Compliance Certification Services Inc. (Tainan Lab.)</b> No.8,Jiucengling, Xinhua Dist., Tainan City 712, Taiwan (R.O.C.)

This device has been tested and found to comply with the stated standard(s), which is (are) required by the Council Directive of 2004/108/EC\* and 2014/30/EU. The test results are indicated in the test report and are applicable only to the tested sample identified in the report.

Note: The above EN basic standards are applied with latest version if customer has no special requirement. \*: Directive 2004/108/EC is repealed with effect from 20 April 2016.

Jeter Wu / Assistant Manager  
Tainan Lab.

Date: March 31, 2016