



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.

**Technical Standard: FCC Part 15 Class B
IC ICES-003**

General Information

Applicant: FSP Group Inc.
No 22, Jianguo East Road, Taoyuan City,
Taiwan 330, R.O.C.

Product Description

EUT Description: Switching Power Adapter

Model Number: FSP120-ABAN2; FSP120-ABBN2; FSP135-ABAN2; FSP135-ABBN2;
FSP120-AAAN2; FSP120-AABN2; FSP096-AHAN2; FSP096-AHBN2;
FSP120-AFAN2; FSP120-AFBN2; FSP120-AWAN2; FSP120-AWBN2;
PUP121-12yyyyy; PUP121S-12yyyyy; PUP121-13-2yyyyy;
PUP121-14yyyyy; PUP121-18yyyyy; PUP121-19yyyyy;
PUP121S-13-2yyyyy; PUP121S-14yyyyy; PUP121S-18yyyyy;
PUP121S-19yyyyy; PUP136-13-2yyyyy; PUP136S-13-2yyyyy
(y=0-9, A-Z, hyphen or blank); FSP120-AFAN2-H3

Measurement Facilities

Xindian Lab.: **Compliance Certification Services Inc.**
No.163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, 23151 Taiwan.
Tel: +886-2-22170894 / Fax: +886-2-22171029

This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T160718D04-D

Sam Hu / Assistant Manager

Date: July 26, 2016