


VERIFICATION OF COMPLIANCE

Issue Date: January 11, 2022
Applicant: FSP Group Inc.
Address: No. 22, Jianguo E. Road, Taoyuan 330 Taiwan
Manufacturer: N/A
Address: N/A
Contact Information: Web: N/A
TEL#: N/A
E-mail#: N/A
Product: Switching Power Supply
Brand Name/Trade Mark: **FSP** ;  ; **TRUMPower**
Model/Type: FSP600M-70PB
Added Model(s): FSP700M-70PB; TMPC600-70G; TMPC700-70G
Applicable Standards: IEC 60601-1-2: 2014
EN 60601-1-2: 2015
CISPR 11: 2015 + A1: 2016 (Ed 6.1)
EN 55011: 2016 / A11: 2020 (Group 1, Class B)
IEC 61000-4-2: 2008; EN 61000-4-2: 2009
IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010;
EN 61000-4-3: 2006 + A1: 2008 + A2: 2010
IEC 61000-4-4: 2012; EN 61000-4-4: 2012
IEC 61000-4-5: 2014 + A1: 2017;
EN 61000-4-5: 2014 + A1: 2017
IEC 61000-4-6: 2013 + COR1: 2015;
EN 61000-4-6: 2014 + AC: 2015
IEC 61000-4-8: 2009; EN 61000-4-8: 2010
IEC 61000-4-11: 2004 + A1: 2017;
EN 61000-4-11: 2004 + A1: 2017
IEC 61000-3-2: 2018 / A1: 2020; EN IEC 61000-3-2: 2019 / A1: 2021
IEC 61000-3-3: 2013 / A1: 2017; EN 61000-3-3: 2013 / A1: 2019


Test Laboratory: Compliance Certification Services Inc.
Xindian Lab
No. 163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, Taiwan

Test Report No. : TMXD2112001733DE, dated on January 11, 2022

Conclusion: Based upon a review of the Test Report(s), the tested sample of the product mentioned above is deemed to comply with the requirements of the above standards

Note: This verification is only valid for the product and configuration described and in conjunction with the test report as detailed above.

Authorised Signatory:



Compliance Certification Services Inc.
Jason Lee
Section Manager

FM-023A-R01