

VERIFICATION OF COMPLIANCE

Issue Date: June 14, 2024
Applicant: FSP Group Inc.
Address: No 22, Jianguo East Road, Taoyuan City, Taiwan 330, R.O.C.
Manufacturer: FSP Group Inc.
Address: No 22, Jianguo East Road, Taoyuan City, Taiwan 330, R.O.C.
Product: Switching Power Supply
Brand Name/Trade Mark: FSP
Model/Type: FSP300M-50PE; FSP400M-50PE; FSP300M-50PEXXX("X"=0-9, A-Z, hyphen or blank.); FSP400M-50PEXXX("X"=0-9, A-Z, hyphen or blank.)
Applicable Standards: IEC 60601-1-2: 2014 + A1: 2020
 EN 60601-1-2: 2015 + A1: 2021
 CISPR 11: 2015 + A1: 2016 + A2: 2019 (Ed 6.2)
 EN 55011: 2016 + A1: 2017 + A11: 2020 + A2: 2021 (Group 1)
 IEC 61000-3-2: 2018 + A1: 2020 (Ed. 5.1)
 EN IEC 61000-3-2: 2019 + A1: 2021
 IEC 61000-3-3: 2013 + AMD1: 2017 + AMD2: 2021 + COR1: 2022
 EN 61000-3-3: 2013 + A1: 2019 + A2: 2021 + AC: 2022
 IEC 61000-4-2: 2008; EN 61000-4-2: 2009
 IEC 61000-4-3: 2020 (Ed. 4.0); EN IEC 61000-4-3: 2020
 IEC 61000-4-4: 2012; EN 61000-4-4: 2012
 IEC 61000-4-5: 2014 + A1: 2017; EN 61000-4-5: 2014 + A1: 2017
 IEC 61000-4-6: 2013 + COR1: 2015; EN 61000-4-6: 2014 + AC: 2015
 IEC 61000-4-8: 2009; EN 61000-4-8: 2010
 IEC 61000-4-11: 2020 + COR1: 2020 + COR2: 2022 (Ed. 3.0);
 EN IEC 61000-4-11: 2020 + AC: 2022-10
 IEC 61000-4-39: 2017; EN 61000-4-39: 2017
 EN 60601-1-2: 2015 + A1: 2021 Subclause 8.11 (IEC 61000-4-39 for §8.10)
Test Laboratory: Compliance Certification Services Inc.
 Xindian Lab
 No. 163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, Taiwan
Test Report No.: TMXD2405001749DE, dated on June 14, 2024
Conclusion: Based upon a review of the Test Report(s), the tested sample of the product mentioned above is deemed to comply with the requirements of the above standards.
Note: This verification is only valid for the product and configuration described and in conjunction with the test report as detailed above.

Authorised Signatory:



Compliance Certification Services Inc.
Jason Lee
Section Manager