

# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20160115-E211696  
**Report Reference** E211696-20160112  
**Issue Date** 2016-JANUARY-15

**Issued to:** FSP TECHNOLOGY INC  
22 JIANGUO E RD  
TAOYUAN CITY  
330 TAIWAN

**This is to certify that  
representative samples of**

COMPONENT - POWER SUPPLIES, MEDICAL AND  
DENTAL

See addendum page

Have been investigated by UL in accordance with the  
Standard(s) indicated on this Certificate.

**Standard(s) for Safety:**

ANSI/AAMI ES60601-1, Medical Electrical Equipment - Part  
1: General Requirements for Basic Safety and Essential  
Performance  
CAN/CSA-C22.2 No. 60601-1, Medical Electrical  
Equipment - Part 1: General Requirements for Basic Safety  
and Essential Performance

**Additional Information:**

See the UL Online Certifications Directory at  
[www.ul.com/database](http://www.ul.com/database) for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's  
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance  
capabilities and are intended for use as components of complete equipment submitted for investigation rather  
than for direct separate installation in the field. The final acceptance of the component is dependent upon its  
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program  
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please  
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



# CERTIFICATE OF COMPLIANCE

**Certificate Number** 20160115-E211696  
**Report Reference** E211696-20160112  
**Issue Date** 2016-JANUARY-15

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Switching Power Adapter,

- 1) FSP065-DHCM1, FSP065-DSCM1, FSP065-DGCM1, FSP065-DDCM1, FSP065-DBCM1, FSP065-DCCM1, FSP065-DACM1;
- 2) PMP65SF-12yyyyy, PMP65SF-12-2yyyyy, PMP65SF-13yyyyy, PMP65SF-13-1yyyyy, PMP65SF-13-2yyyyy, PMP65SF-13-3yyyyy, PMP65SF-14yyyyy (y=0-9, A-Z, hyphen or blank)(Variable of model designation used for marketing purposes only, not design relevant)



Bruce Mahrenholz, Director North American Certification Program  
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

