



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company and for below described product, based on

**Technical Standard : MDD DIRECTIVE 93/42/EEC
(EN 60601-1-2)**

General Information

Applicant : FSP Group Inc.
Address of Applicant : No 22, Jianguo East Road, Taoyuan City,
Taiwan 330, R.O.C.

Product Description

Product Name : Switching Power Adapter
Brand Name : FSP; PROTEK
Model Number : FSP150M-AHA; FSP150M-AGA; FSP150M-ABA; FSP150M-AAA;
: PMP150A-12yyyyyy; PMP150A-13yyyyyy; PMP150A-13-2yyyyyy;
PMP150A-14yyyyyy ("y"=0-9, A-Z, hyphen or blank)

Measurement Standard

IEC 60601-1-2: 2014
EN 60601-1-2: 2015
CISPR 11: 2009 + A1: 2010
EN 55011: 2009 + A1: 2010 (Group 1, Class B)
IEC 61000-3-2: 2014; EN 61000-3-2: 2014
IEC 61000-3-3: 2013; EN 61000-3-3: 2013
IEC 61000-4-2: 2008; EN 61000-4-2: 2009; IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010;
EN 61000-4-3: 2006 + A1: 2008 + A2: 2010; IEC 61000-4-4: 2012; EN 61000-4-4: 2012;
IEC 61000-4-5: 2014; EN 61000-4-5: 2014; IEC 61000-4-6: 2013; EN 61000-4-6: 2014;
IEC 61000-4-8: 2009; EN 61000-4-8: 2010; IEC 61000-4-11: 2004; EN 61000-4-11: 2004

Measurement Facilities

Company Name : **Compliance Certification Services Inc.**
Test Laboratory : Xindian Lab.
Address of Test Lab. : No.163-1, Jhongscheng Rd., Xindian Dist., New Taipei City, 23151 Taiwan.

This device has been tested and found to be in compliance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report with the number: T170920D07-E

The test results shown in this report are applicable only to the investigated sample identified in this report.

Sam Hu / Assistant Manager
Date: September 28, 2017