



## VERIFICATION OF COMPLIANCE

*This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.*

**Technical Standard: EMC DIRECTIVE 93/42/EEC  
(EN 60601-1-2)**

### General Information

Applicant: FSP Group Inc.  
No 22, Jianguo East Road, Taoyuan City,  
Taiwan 330, R.O.C.

### Product Description

EUT Description: Switching Power Supply  
FSP035M-B24-05; FSP045M-B24-12; FSP045M-B24-15; FSP045M-B24-18;  
Model Number: FSP045M-B24-24; THC45-05; THC45-12; THC45-15; THC45-18; THC45-24; PM45-10A;  
PM45-12A; PM45-13A; PM45-13-1A; PM45-14A

### Measurement Standard

IEC 60601-1-2: 2014  
EN 60601-1-2: 2015  
CISPR 11: 2009 + A1: 2010  
EN 55011: 2009 + A1: 2010 (Group 1, Class B)  
IEC 61000-3-2: 2014; EN 61000-3-2: 2014  
IEC 61000-3-3: 2013; EN 61000-3-3: 2013  
IEC 61000-4-2: 2008; EN 61000-4-2: 2009; IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010;  
EN 61000-4-3: 2006 + A1: 2008 + A2: 2010; IEC 61000-4-4: 2012; EN 61000-4-4: 2012;  
IEC 61000-4-5: 2014; EN 61000-4-5: 2014; IEC 61000-4-6: 2013; EN 61000-4-6: 2014;  
IEC 61000-4-8: 2009; EN 61000-4-8: 2010; IEC 61000-4-11: 2004; EN 61000-4-11: 2004

### Measurement Facilities

Xindian Lab.: **Compliance Certification Services Inc.**  
No.163-1, Zhongsheng Rd., Xindian Dist., New Taipei City, 23151 Taiwan.  
Tel: +886-2-22170894 / Fax: +886-2-22171029

*This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T160314D01-E*

*Sam Hu / Assistant Manager*

*Date: July 19, 2016*