



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the product designated below.

Product **Switching Power Supply**

Model **PM100-xyyyy**
(x=10, 12, 13, 13-1, 14, 15, 17, 18 ; y=hyphen, A-Z, 0-9 or blank);
FSP100-xM1
(x=1k00, 1k20, 1k30, 1k31, 1k40, 1k50, 1k70, 1k80)

Applicant **FSP Group Inc.**
2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C.

Manufacturer **(1)FSP Technology Inc. Kaohsiung Branch.**
2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C.
(2)Protek Electronic (China) Corp.
No. 3, Yong Tai 1st Street , North Area Industry Tian Mei Huang Jiang,
Dongguan, Guangdong 523752 China
(3)Zhonghan Electronics (Shenzhen) Co., Ltd.
Building A2 in Section A, Building 4,6,7 in Section C, JuYuan Industrial Zone,
TangWei Village, FuYong Town, BaoAn Dist., ShenZhen City, Guangdong
518103 China

Applicable Standard(s) **EN 55011: 2009+A1: 2010**
EN 61000-3-2: 2006+A1: 2009+A2: 2009, EN 61000-3-3: 2013
EN 60601-1-2: 2014
IEC 61000-4-2: 2008, IEC 61000-4-3: 2010,
IEC 61000-4-4: 2012, IEC 61000-4-5: 2014,
IEC 61000-4-6: 2013, IEC 61000-4-8: 2009,
IEC 61000-4-11: 2004

Reference No. **T150521N06-E**

Test Laboratory **Compliance Certification Services Inc. (Tainan Lab.)**
No.8,Jiucengling, Xinhua Dist., Tainan City 712, Taiwan (R.O.C.)

This device has been tested and found to comply with the stated standard(s), which is (are) required by the Council Directive of 93/42/EEC, 2004/108/EC* and 2014/30/EU, Amended by 92/31/EEC, 2006/95/EC & 98/13/EC. The test results are indicated in the test report and are applicable only to the tested sample identified in the report.

Note: The above EN basic standards are applied with latest version if customer has no special requirement. *: Directive 2004/108/EC is repealed with effect from 20 April 2016.

Jeter Wu / Assistant Manager
Tainan Lab.

Date: June 29, 2015



程智科技股份有限公司
Compliance Certification Services Inc.