




VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the product designated below.

Product **Switching power supply**
Model **PM301-xyzf (x=12, 13, 13-2, 14, 16, 17 and 18, y=A, B or C, z= blank or “-12”, f =blank or “SP”)**
FSP300M-K36-xyzf (x=12,15,19,24,30,36 and 48, y=A,B or C, z= blank or “H”, f= blank or “SP”)
Brand **FSP ; PROTEK POWER**
Applicant **FSP Group Inc.**
2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C.
Manufacturer **(1) FSP Technology Inc. Kaohsiung Branch.**
2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C.
(2) Protek Electronic (China) Corp.
No. 3, Yong Tai 1st Street , North Area Industry Tian Mei Huang Jiang,
Dongguan, Guangdong 523752 China
(3) Zhonghan Electronics (Shenzhen) Co., Ltd.
Building A2 in Section A, Building 4,6,7 in Section C, JuYuan Industrial
Zone, TangWei Village, FuYong Town, BaoAn Dist., ShenZhen City,
Guangdong 518103 China
Applicable Standard(s) **EN 55011: 2009 + A1: 2010 (Group 1, Class B)**
CISPR 11: 2011
EN 61000-3-2: 2014; IEC 61000-3-2: 2009
EN 61000-3-3: 2013; IEC 61000-3-3: 2013
EN 60601-1-2: 2007 + AC: 2010
IEC 60601-1-2: 2014 (Edition 4)
IEC 61000-4-2: 2008; EN 61000-4-2: 2009
IEC 61000-4-3: 2010; EN 61000-4-3: 2006 + A1: 2008 + A2: 2010
IEC 61000-4-4: 2012; EN 61000-4-4: 2012
IEC 61000-4-5: 2014; EN 61000-4-5: 2014
IEC 61000-4-6: 2013; EN 61000-4-6: 2014
IEC 61000-4-8: 2009; EN 61000-4-8: 2010
IEC 61000-4-11: 2004; EN 61000-4-11: 2004
Reference No. T151216N16-E
Test Laboratory **Compliance Certification Services Inc. (Tainan Lab.)**
No.8,Jiucengling, Xinhua Dist., Tainan City 712, Taiwan (R.O.C.)

This device has been tested and found to comply with the stated standard(s), which is (are) required by the Council Directive of 93/42/EEC, 2004/108/EC* and 2014/30/EU, Amended by 92/31/EEC, 2006/95/EC & 98/13/EC. The test results are indicated in the test report and are applicable only to the tested sample identified in the report.

Note: The above EN basic standards are applied with latest version if customer has no special requirement. *: Directive 2004/108/EC is repealed with effect from 20 April 2016.


Jeter Wu / Assistant Manager

Tainan Lab.

Date: February 19, 2016