

NOTICE OF COMPLETION
AND
AUTHORIZATION TO APPLY THE UL MARK



02/12/2015

FSP GROUP INC
CAITER LEE
2-3 E 3RD ST PO BOX 35-25 N E P Z
KAOHSIUNG, 811 TW

Our Reference: File E178020, Vol.2 S4 Project Number 4786784123
Your Reference:
Project Scope: UL/cUL: 1.change model name to FSP300M-K36-xyzf series, 2.change Y-capacitor CY5, CY6 (E178020-V2S4)

Dear Mr. Lee:

Congratulations! UL's investigation of your product(s) has been completed under the above Reference Number and the product was determined to comply with the applicable requirements. This letter temporarily supplements the UL Follow-Up Services Procedure and serves as authorization to apply the UL Mark at authorized factories under UL's Follow-Up Service Program. To provide your manufacturer(s) with the intended authorization to use the UL Mark, you must send a copy of this notice to each manufacturing location currently authorized under File E178020, Vol. 2 S4 and *including any special instructions as indicated in the addendum to this letter.*

Records in the Follow-Up Services Procedure covering the product are now being prepared and will be sent in the near future. Until then, this letter authorizes application of the UL Mark for 90 days from the date indicated above.

Additional requirements related to your responsibilities as the Applicant can be found in the document "Applicant responsibilities related to Early Authorizations" that can be found at the following web-site:
<http://www.ul.com/EAResponsibilities>

Any information and documentation provided to you involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

We are excited you are now able to apply the UL Mark to your products and appreciate your business. Feel free to contact me or any of our Customer Service representatives if you have any questions.

Very truly yours,

Neo Chang
Engineer
Health Sciences
UL Taiwan Co., Ltd

Reviewed by:

Bruce Mahrenholz
CPO Director
Certification Program Office
UL LLC

CERTIFICATE OF COMPLIANCE

Certificate Number 20150212-E178020
Report Reference E178020-20141219
Issue Date 2015-FEBRUARY-12

Issued to: FSP GROUP INC
2-3 E 3RD ST
PO BOX 35-25
N E P Z
KAOHSIUNG 811 TAIWAN

**This is to certify that
representative samples of**

COMPONENT - POWER SUPPLIES, MEDICAL AND
DENTAL

See addendum page

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety:

ANSI/AAMI ES60601-1, Medical Electrical Equipment - Part
1: General Requirements for Basic Safety and Essential
Performance
CAN/CSA-C22.2 No. 60601-1, Medical Electrical
Equipment - Part 1: General Requirements for Basic Safety
and Essential Performance

Additional Information:

See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance
capabilities and are intended for use as components of complete equipment submitted for investigation rather
than for direct separate installation in the field. The final acceptance of the component is dependent upon its
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at www.ul.com/contactus



CERTIFICATE OF COMPLIANCE

Certificate Number	20150212-E178020
Report Reference	E178020-20141219
Issue Date	2015-FEBRUARY-12

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Switching Power Supply

PM301-xyzf

x = 12, 13, 13-2, 14, 16, 17 or 18

y = A, B or C

z = blank or -12

f = blank or -SP

FSP300M-K36-xyzf

x = 12, 15, 19, 24, 30, 36 or 48

y = A, B or C

z = blank or H

f = blank or -SP

Bruce Mahrenholz, Assistant Chief Engineer, Global Inspection and Field Services

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at www.ul.com/contactus

