

CERTIFICATE OF COMPLIANCE

Certificate Number 20160330-E178020
Report Reference E178020-V2-S10
Issue Date 2016-MARCH-30

Issued to: FSP GROUP INC
2-3 E 3RD ST, PO BOX 35-25
N E P Z, KAOHSIUNG
811 TAIWAN

This is to certify that representative samples of Power Supplies, Medical and Dental
See next page for models.

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.


Standard(s) for Safety: ANSI/AAMI ES60601-1, Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance
CAN/CSA-C22.2 No. 60601-1:14, Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance

Additional Information: See the UL Online Certifications Directory at www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance capabilities and are intended for use as components of complete equipment submitted for investigation rather than for direct separate installation in the field. The final acceptance of the component is dependent upon its installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>



CERTIFICATE OF COMPLIANCE

Certificate Number 20160330-E178020
Report Reference E178020-V2-S10
Issue Date 2016-MARCH-30

This is to certify that representative samples of the product as specified on this certificate were tested according to the current UL requirements.

Switching Power Supply.

Models:


PM400-xy (x = 12, 13, 13-1, 14, 15, 17, 18; y = B or C)

PMP400-x-S (x=13-1, 14, 15, 17, 18)

FSP400-1KzM1 (z = 20, 30, 31, 40, 50, 70 or 80)

FSP400-1KwM1C (w = 20, 30, 31, 40, 50, 70 or 80)

FSP400-KxAM1 (x = D, A, K, E or F)



Bruce Mahrenholz, Director North American Certification Program

UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

