

**AUDIX TECHNOLOGY CORPORATION****EMC DEPARTMENT**

No. 53-11, Dingfu, Linkou Dist.,

New Taipei City 244, Taiwan.

Tel: 886-2-2609-9301~2 Fax: 886-2-2609-9303

HEAD OFFICE

No. 8, Lane 120, Sec. 1, Nei Hu Rd.,

Taipei, Taiwan.

Tel: 886-2-2659-4900 Fax: 886-2-2659-4833

Technical Compliance Statement


No. EM-E160240

The following products has been tested by us with the listed standards and found in compliance with the council EMC directive 93/42/EEC. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

Submittor : FSP Group Inc.**No. 22, Jianguo E., Rd., Taoyuan City, Taiwan****Product : Switching Power Supply****M/N (1)FSP250-60MPC (2)FSP220-60MPC****(3)SPI250F4BMB (4)SPI220F4BMB****Brand (1)~(2)FSP (3)~(4)SPI****Test Standards**

EN 55011:2009 +A1:2010	Limits and methods of measurement of radio disturbance characteristics of information technology equipment
EN 60601-1-2:2007 +AC:2010 IEC 60601-1-2:2014	Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests
CISPR 11:2009 +A1:2010	Limits and methods of measurement of radio disturbance characteristics of information technology equipment
EN 61000-3-2:2014 IEC 61000-3-2:2014	Part 3 : Limits-Section 2 : Limits for harmonic current emission (equipment input current <=16A per phase)
EN 61000-3-3:2013 IEC 61000-3-3:2013	Part 3 : Limits-Section 3 : Limitation of voltage fluctuations and flicker in low-voltage supply systems for equipment with rated current <=16A
EN 61000-4-2:2009 IEC 61000-4-2:2008	Electrostatic discharge immunity test
EN 61000-4-3:2006 +A1:2008 +A2:2010 IEC 61000-4-3:2010	Radiated, radio-frequency electromagnetic field immunity test
EN 61000-4-4:2012 IEC 61000-4-4:2012	Electrical fast transient / burst immunity test
EN 61000-4-5:2014 IEC61000-4-5:2014	Surge immunity test
EN 61000-4-6:2014 IEC 61000-4-6:2013	Immunity to conducted disturbances, induced by radio-frequency fields
EN 61000-4-8:2010 IEC 61000-4-8:2009	Power Frequency Magnetic field immunity test
EN 61000-4-11:2004 IEC 61000-4-11:2004	Voltage dips, short interruptions and voltage variations immunity test




 Alex Deng/Deputy Manager
 Date: 2016. 04. 22

The verification is based on a single evaluation of one sample of above mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. logo.