

# Technical Compliance Statement

## EMC Test Report

For the following information

Ref. File No.: C1M2305071

Product : Switching Power Supply  
Model Name : (1)FSP250-60MPC (2)FSP220-60MPC (3)SPI250F4BMB  
(4)SPI220F4BMB  
Brand Name : (1)~(2)FSP (3)~(4)SPI  
Applicant : FSP Group Inc.  
Standards :

EN 60601-1-2:2015 +A1:2021 (IEC 60601-1-2:2014 +AMD:2020)(Edition 4.1)

BS EN 60601-1-2:2015 +A1:2021

Emission: CISPR 11:2015 +A1:2016 +A2:2019, Class B

EN 55011:2016 +A11:2020

EN IEC 61000-3-2:2019+A1:2021 (IEC 61000-3-2:2018 +A1:2020),

BS EN IEC 61000-3-2:2019+A1:2021 and

EN 61000-3-3:2013 +A1:2019 (IEC 61000-3-3:2013 +A1:2017),

EN 61000-3-3:2013 +A1:2019 +A2:2021

(IEC 61000-3-3:2013 +A1:2019 +A2:2021)

BS EN 61000-3-3:2013 +A1:2019,

BS EN 61000-3-3:2013 +A1:2019 +A2:2021

Immunity: IEC 61000-4-2:2008, IEC 61000-4-3:2020


IEC 61000-4-4:2012, IEC 61000-4-5:2014+A1:2017,

IEC 61000-4-6:2013, IEC 61000-4-8:2009,

IEC 61000-4-11:2020, IEC 61000-4-39:2017

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the Regulation (EU) 2017/745 concerning medical device. The test data and results are included in the test report **EM-E160240B**.

Signature



Tony Lee/Deputy Manager

Date: 2023. 05. 24

Test Laboratory:

Audix Technology Corporation, EMC Department

TAF Accreditation No.: 1724

Web Site: [www.audixtech.com](http://www.audixtech.com)

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.