

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OCCB TEST CERTIFICATE
CERTIFICAT D'ESSAI OCProduct
Produit

Switching Power Supply

Name and address of the applicant
Nom et adresse du demandeurFSP Group Inc.
No. 22, Jianguo E. Rd.
330 Taoyuan City, TaiwanName and address of the manufacturer
Nom et adresse du fabricantFSP Group Inc.
No. 22, Jianguo E. Rd.
330 Taoyuan City, TaiwanName and address of the factory
Nom et adresse de l'usine Additional Information page 2Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} pageRatings and principal characteristics
Valeurs nominales et caractéristiques principalesAC 100-240V; 60-50Hz; 10-5A(for FSP700-90MPB,FSP700-92MPB);
8-4A(for FSP600-90MPB, FSP600-92MPB); 13-8A(for FSP700-90MPB
); 11-7A(for FSP600-90MPB); Class I;
for output ratings refer to test report page 9Trade mark (if any)
Marque de fabrique (si elle existe)

FSP

Model/type Ref.
Ref. de typeFSPxxx-yyMPB
(xxx = 600 or 700 for output rating; yy = 90 or 92 to denote
cooling option (fan size and location))Additional information (if necessary may also be
reported on page 2)
Les Information complémentaire (si nécessaire,
peuvent être indiqués sur la 2^{ème} page)Models differ model name, DC fan, AC outlet, input rating,
output ratings, and the rating of mains fuse, ripple
capacitor 1C1, bridge diodes BD1, BD2, transistors 1Q10,
1Q12; for details see test report

PUBLICATION

EDITION

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à laIEC 60601-1:2005+A1
for national deviations see test reportAs shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue une partie de ce Certificat

11045730 001

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de CertificationTÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com

Date: 13.05.2016

Signature:


Dipl.-Ing. W. Hsu

1. Zhonghan Electronics (ShenZhen) Co., Ltd., Building A2 in Section A Building 4,6,7 in Section C, JuYuan Industrial Zone, TangWei Village, FuYong Town, BaoAn Dist., ShenZhen City, Guangdong, China
2. Wuxi ZhongHan Technology Co., Ltd. Block 106-E, Wuxi Nation HI-TECH Industrial, Development Zone, Wuxi City, Jiangsu, Province, China
3. ZHONGHAN SCIENCE & TECH CO.LTD, Building 9, Section B JuYuan Industrial Zone, TangWei Village, FuYong Town, BaoAn District ShenZhen City, Guangdong, China
4. ShenZhen HuiLi Electronics CO., LTD Block C, Building 4,6,7,8,9,10,11, County 73, Xin'an, Bao'an, Shenzhen, Guangdong, China
5. ShenZhen HuiLi Electronics CO., LTD. Block C, Building 4,6,7,8,9,10,11, County 73, Xin'an, Bao'an, (Block A, Building 3, No. 5185, YeeYuan Rd., County 74, Bao'an), Shenzhen, Guangdong, China
6. Wuxi SPI Technology Co., Ltd. No. 96, Xinmei Road, New District Wuxi City, Jiangsu China

Additional information (if necessary)
Information complémentaire (si nécessaire)



Date: 13.05.2016

Signature:


Dipl.-Ing. W. Hsu