



## VERIFICATION OF COMPLIANCE

**This Verification of Compliance is hereby issued to the product designated below.**

Product	Switching power supply
Model	PM150-xAy(x=12, 13, 13-1, 14, 16, 16-1, 17 and 18, y= blank or “-SP”); FSP150M-K24-xy(x=12,15,18,24,30,36 and 48, y= blank or “-SP”)
Brand Name	PROTEK POWER; FSP
Applicant	FSP Group Inc. 2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C.
Manufacturer	(1) FSP Technology Inc. Kaohsiung Branch. 2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C. (2) Protek Electronic (China) Corp. No. 3, Yong Tai 1st Street , North Area Industry Tian Mei Huang Jiang, Dongguan, Guangdong 523752 China (3) Zhonghan Electronics (Shenzhen) Co., Ltd. Building A2 in Section A, Building 4,6,7 in Section C, JuYuan Industrial Zone, TangWei Village,FuYong Town, BaoAn Dist., ShenZhen City, Guangdong 518103 China
Applicable Standard(s)	IEC 60601-1-2: 2014 (Edition 4) EN 55011: 2009 + A1: 2010 (Class B) EN 60601-1-2: 2007 + AC: 2010 EN 60601-1-2: 2015 EN 61000-3-2: 2014, EN 61000-3-3: 2013, IEC 61000-4-2: 2008, IEC 61000-4-3: 2010, IEC 61000-4-4: 2012, IEC 61000-4-5: 2014, IEC 61000-4-6: 2013, IEC 61000-4-8: 2009, IEC 61000-4-11: 2004
Reference No.	T161107N02-E
Test Laboratory	Compliance Certification Services Inc. (Tainan Lab.) No.8, Jiucengling, Xinhua Dist., Tainan City 712, Taiwan (R.O.C.)

**This device has been tested and found to comply with the stated standard(s), which is(are) required by the Council Medical Directive & Electromagnetic Compatibility Directive (93/42/EEC ; 2014/30/EU ). The test results are indicated in the test report and are applicable only to the tested sample identified in the report.**

Jeter Wu / Assistant Manager  
Tainan Lab.  
Date: December 15, 2016