

# Certificate



Certificate no.

TA 50394533 01

**License Holder:**

FSP Group Inc.  
No. 22, Jianguo E. Road  
Taoyuan 330  
Taiwan

**Manufacturing Plant:**

ShenZhen HuiLi Electronics CO., LTD  
Block C, Building 4,6,7,8,9,10,11,  
County 73, Xin'an, Bao'an,  
Shenzhen, Guangdong  
P. R. China

**Test report no.:** ZTW1-SCL 50105417 001

**Client Reference:** ZD-18092017-FSP500M/

**Tested to:** IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

**Certified Product:** (Switching Power Supply)

**License Fee - Units**

Type Designation	: FSP500M-80PA (FSP)	10
Rated Voltage	: AC 100-240V, 60-50Hz	
Rated Current	: 8-4A	
Output Voltage	: see constructional dataform	
Output Current	: see constructional dataform	
max. Ambient Temperature:	+50°C	
Protection Class	: I	

This certificate is valid for maximum 5 years.

Appendix: 1

10

**Licensed Test mark:**



**TÜV Rheinland Taiwan Ltd**

Signature

**Date of Issue**  
(day/mo/yr)

08/12/2017



**Dipl.-Ing. (FH) A. Klinker**

# Certificate



Certificate no.

TA 50394533 02

**License Holder:**

FSP Group Inc.  
No. 22, Jianguo E. Road  
Taoyuan 330  
Taiwan

**Manufacturing Plant:**

ShenZhen HuiLi Electronics CO., LTD  
Block C, Building 4,6,7,8,9,10,11,  
County 73, Xin'an, Bao'an,  
Shenzhen, Guangdong  
P. R. China

**Test report no.:** ZTW1-SCL 50105417 001

**Client Reference:** ZD-18092017-FSP500M/

**Tested to:** IEC 60601-1:2005+A1  
EN 60601-1:2006+A11+A1+A12

**Certified Product:** (Switching Power Supply)

**License Fee - Units**

as page 01

Addition

Type Designation	: FSP500M-80UA (FSP);	1
	FSP400M-80PA (FSP)	1

This certificate is valid for maximum 5 years.

Appendix: 1

2

**Licensed Test mark:**



**TÜV Rheinland Taiwan Ltd**

Signature

**Date of Issue**

(day/mo/yr)

08/12/2017



Dipl.-Ing. (FH) A. Klinker