



## VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company and for below described product, based on

**Technical Standard : MDD DIRECTIVE 93/42/EEC  
(EN 60601-1-2)**

### General Information

Applicant : FSP Group Inc.  
Address of Applicant : No 22, Jianguo East Road, Taoyuan City,  
Taiwan 330, R.O.C.

### Product Description

Product Name : Switching Power Supply  
Model Number : FSP500M-80PA; FSP500M-80UA; FSP400M-80PA

### Measurement Standard

IEC 60601-1-2: 2014  
EN 60601-1-2: 2015  
CISPR 11: 2009 + A1: 2010  
EN 55011: 2009 + A1: 2010 (Group 1, Class B)  
IEC 61000-3-2: 2014; EN 61000-3-2: 2014  
IEC 61000-3-3: 2013; EN 61000-3-3: 2013  
IEC 61000-4-2: 2008; EN 61000-4-2: 2009; IEC 61000-4-3: 2006 + A1: 2007 + A2: 2010;  
EN 61000-4-3: 2006 + A1: 2008 + A2: 2010; IEC 61000-4-4: 2012; EN 61000-4-4: 2012;  
IEC 61000-4-5: 2014; EN 61000-4-5: 2014; IEC 61000-4-6: 2013; EN 61000-4-6: 2014;  
IEC 61000-4-8: 2009; EN 61000-4-8: 2010; IEC 61000-4-11: 2004; EN 61000-4-11: 2004

### Measurement Facilities

Company Name : **Compliance Certification Services Inc.**  
Test Laboratory : Xindian Lab.  
Address of Test Lab. : No.163-1, Jhongsheng Rd., Xindian Dist., New Taipei City, 23151 Taiwan.

This device has been tested and found to be in compliance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report with the number: T171016D03-E

The test results shown in this report are applicable only to the investigated sample identified in this report.

Sam Hu / Assistant Manager  
Date: October 23, 2017