



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the product designated below.

Product	Switching power supply
Model	a) PMP30x-yzzzzzzz (z x7) series; x=M (Interchangeable), NP (Fixed North American, Non-Polarization), E (Fixed European) or CN(Fixed China) for AC plug' s style; y=10,10-2, 11, 11-1, 12, 13, 13-1, 14, 16 or 18 for different output voltage; z=0-9, A-Z, hyphen or blank for marking propose b) FSPnM-Kxwzzzzzzz(z x7) series (n=020, 027, 030; x= P, R, S, H, G, D, A, L or F; w= M, U , E or N; z=0-9, A-Z, hyphen or blank); n=020, 027, 030 for output power; x= P, R, S, H, G, D, A, L or F for different output voltage; w= M (Interchangeable), U (Fixed North American, Non-Polarization), E (Fixed European) or N (Fixed China) for AC plug' s style; z=0-9, A-Z, hyphen or blank for marking propose
Brand	PROTEK POWER & FSP
Applicant	FSP Group Inc. 2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan
Manufacturer	(1) FSP Technology Inc. Kaohsiung Branch. 2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan (2) Protek Electronic (China) Corp. No. 3, Yong Tai 1st Street , North Area Industry Tian Mei Huang Jiang, Dongguan, Guangdong 523752 China (3) Zhonghan Electronics (Shenzhen) Co., Ltd. Building A2 in Section A, Building 4,6,7 in Section C, JuYuan Industrial Zone, TangWei Village, FuYong Town, BaoAn Dist., ShenZhen City, Guangdong 518103 China
Applicable Standard(s)	IEC 60601-1-2: 2014 (Edition 4) EN 55011: 2009 + A1: 2010 (Class B) EN 60601-1-2: 2015 EN 61000-3-2: 2014, EN 61000-3-3: 2013, IEC 61000-4-2: 2008, IEC 61000-4-3: 2010, IEC 61000-4-4: 2012, IEC 61000-4-5: 2014, IEC 61000-4-6: 2013, IEC 61000-4-8: 2009, IEC 61000-4-11: 2004+A1: 2017
Reference No.	T190313N03-E
Test Laboratory	Compliance Certification Services Inc. (Tainan Lab.) No.8,Jiucengling, Xinhua Dist., Tainan City 712, Taiwan (R.O.C.)

This device has been tested and found to comply with the stated standard(s), which is (are) required by the Council Medical Directive & Electromagnetic Compatibility Directive (93/42/EEC ; 2014/30/EU). The test results are indicated in the test report and are applicable only to the tested sample identified in the report.

Jeter Wu / Assistant Manager
Tainan Lab.

Date: April 16, 2019

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at www.sgs.com/terms_and_conditions.htm and for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.