



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the product designated below.

Product	Switching power supply
Model	PMP250f-xyyyy(f="F"or blank, x=12, 13, 13-1, 13-2, 13-3, 14, 15, 16, 16-1, 17, 18 and 19, y=hyphen, A-Z,0-9 or blank) FSP250M-Kxyzzzz (x=H, G, D, B, C, A, L, E, F and W; y=A or D, z=hyphen, A-Z,0-9 or blank)
Brand	PROTEK POWER & FSP
Applicant	FSP Group Inc. 2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C.
Manufacturer	(1) FSP Technology Inc. Kaohsiung Branch. 2-3, East 3rd Street, N.E.P.Z., P.O. Box 35-25, Kaohsiung Taiwan, R.O.C. (2) Protek Electronic (China) Corp. No. 3, Yong Tai 1st Street , North Area Industry Tian Mei Huang Jiang, Dongguan, Guangdong 523752 China (3) Zhonghan Electronics (Shenzhen) Co., Ltd. Building A2 in Section A, Building 4,6,7 in Section C, JuYuan Industrial Zone, TangWei Village, FuYong Town, BaoAn Dist., ShenZhen City, Guangdong 518103 China
Applicable Standard(s)	IEC 60601-1-2: 2014 (Edition 4) EN 55011: 2009 + A1: 2010 (Class B) EN 60601-1-2: 2015 EN 61000-3-2: 2014, EN 61000-3-3: 2013 IEC 61000-4-2: 2008, IEC 61000-4-3: 2010,0 IEC 61000-4-4: 2012, IEC 61000-4-5: 2014,2 IEC 61000-4-6: 2013, IEC 61000-4-8: 2009,0 IEC 61000-4-11: 2004+A1: 2017
Reference No.	T171220N01-E
Test Laboratory	Compliance Certification Services Inc. (Tainan Lab.) No.8,Jiucengling, Xinhua Dist., Tainan City 712, Taiwan (R.O.C.)

This device has been tested and found to comply with the stated standard(s), which is (are) required by the Council Medical Directive & Electromagnetic Compatibility Directive (93/42/EEC ; 2014/30/EU). The test results are indicated in the test report and are applicable only to the tested sample identified in the report.

Jeter Wu / Assistant Manager

Tainan Lab.

Date: April 29, 2019