

Certificate



Certificate no.

TA 50434864 01

License Holder:

FSP Group Inc.
No. 22, Jianguo E. Road
Taoyuan 330
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.: ZTW1- 50202849 001

Client Reference: EH/SPC-1810023-app

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1

Certified Product: (Switching Power Adapter)

License Fee - Units

Type Designation	: FSP250M-KzA (FSP)	10
z stands for	: H, G, D, B, C, A, L, E, F or W	1
Rated Voltage	: AC 100-240V, 47-63Hz	
Rated Current	: 2.9-1.25A	
Outputs	: see constructional dataform	
max. Ambient Temperature	: +40°C with o/p loaded to 100% rated +60°C with o/p loaded to 50% rated	
Protection Class	: I	

This certificate is valid for maximum 5 years.

Appendix: 1

11

Licensed Test mark:



IEC 60601-1
EN 60601-1

www.tuv.com
ID 1419068454

TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. (FH) A. Klinker



Date of Issue
(day/mo/yr)

26/04/2019

Anlage Fertigungsstättenliste
/Attachment List of Factories



TA 50434864 0001

- 1 FSP Technology Inc., Kaohsiung Branch
2-3, East 3rd Street
N.E.P.Z., Kaohsiung 811
Taiwan
- 2 Zhonghan Electronics (Shenzhen)
Co., Ltd.
Building A2 in Section A, Building 4, 6, 7
in Section C, JuYuan Industrial Zone,
TangWei Village, FuHai Town,
BaoAn Dist., ShenZhen City, Guangdong
P. R. China
- 3 Protek Electronics (China) Corp.
No. 3, Yong Tai 1st Street, North
Area, Tian Mei Industrial Zone,
Huang Jiang Town, Dong Guan City,
Guang Dong Province 523752
P. R. China



Dieser Anhang ersetzt den Vorgänger vom/
This annex replaces the previous annex dated

Datum / Date 26.04.2019



Zertifizierungsstelle
/ Certification Body

A handwritten signature in blue ink, appearing to be "A. Klinker".

Dipl.-Ing. (FH) A. Klinker

Certificate



Certificate no.

TA 50434864 02

License Holder:

FSP Group Inc.
No. 22, Jianguo E. Road
Taoyuan 330
Taiwan

Manufacturing Plant:

Refer to latest revision
of the annex list of factories

Test report no.:

ZTW1- 50202849 001

Client Reference: EH/SPC-1810023-app

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12
IEC 60601-1-6:2010+A1
EN 60601-1-6:2010+A1
IEC 62366:2007+A1
EN 62366:2008+A1

Certified Product: (Switching Power Adapter)

License Fee - Units

as page 01

Addition

Type Designation	: PMP250-xyyyyyy (PROTEK POWER)	1
x stands for	: 12, 13, 13-1, 13-2, 13-3, 14, 15, 16, 16-1, 17, 18 or 19	1
y stands for	: 0-9, A-Z, hyphen or blank	1

This certificate is valid for maximum 5 years.

Appendix: 1

3

Licensed Test mark:



TÜV Rheinland Taiwan Ltd
Signature

Dipl.-Ing. (FH) A. Klinker



Date of Issue
(day/mo/yr)
26/04/2019